

#### DIGITAL MULTIMETER

# CD800a

#### **APPLICATIONS AND FEATURES**

This instrument is portable multimeter designed for the measurement of low-voltage circuit.

This is used at small communications equipments, home electric appliances, voltage of lump line, and measurement of various types of batteries.

- •4000 count
- •Sharp contrast LCD with character 17.5mm high is used
- •Solid & protective body cover that can also be used as a tilt stand
- Frequency, capacitance, and duty cycle measurement function

## **SPECIFICATIONS**

	Measuring range		Best accuracy	Resolution	Resolution		
DCV	400m/4/40/400/600V	±(0.7%+3)	0.1mV	DCV: 10M~100MΩ			
ACV	4/40/400/600V		±(1.6%+9)	0.001V	ACV: 10M~11MΩ		
DCA	40m/400mA		±(2.2%+5)	0.01mA			
ACA	CA 40m/400mA		±(2.8%+5)	0.01mA			
Resistance	400/4k/40k/400k/4M/40	MΩ	±(1.5%+5)	0.1Ω			
Capacitance	50n/500n/5µ/50µ/100µF	:	±(5.0%+10)	0.01nF			
Frequency	5Hz~100kHz		$\pm(0.5\%+3)$				
Duty cycle	20%~80%	±(0.5%+5)					
Continuity	Buzzer sounds at between $10\Omega$ and $120\Omega$ Open voltage: approx. 0.4V						
Diode test	Open voltage: approx. 1.5V						
Manaruina	م ما الم	hadtem 7 A					
Measuring method		Δ-Σ method					
Display		4000 count					
Range selection		Auto and manual ranges					
Over display		"O.L" is displayed(except AC/DC 600V ranges)					
Polarity		Automatic selection(only " - " is displayed)					
Battery low warning		Battery( = → mark lights when the internal battery's power is below approx. 2.4V					
Sampling rate		2 times / sec					
Operating temperature / humidity		5~40°C / 5~31°C, 80%RH(Max)					
Storage temperature / humidity		-10°C~50°C, below 70%RH No condensation					
Environmental condition		Altitude 2000m or below, pollution degree II					
Power consumption		7mW at DCV					
Continuous use time		About 150 hours at DCV					
Bandwidth		40∼400Hz					
Fuse / Battery		0.5A/250V 1.5kA φ5.2x20mm / R6P x 2					

H176 x W104 x D46mm / 340g Hand strap, Instruction manual



A battery for monitoring has been installed prior to shipment from the factory. It may be discharged before the expiration of the described battery life. This battery is used to check the functions and performance of the product. Specifications and external appearance of the product described above may be revised for modification without prior notice.

# sanwa

Standard accessories included

### SANWA ELECTRIC INSTRUMENT CO., LTD.

Dempa Bldg, 4-4 Sotokanda 2-Chome, Chiyoda-Ku, Tokyo 101-0021 Japan Tel:+81-3-3251-0941 Fax:+81-3-3256-9740

www.sanwa-meter.co.jp

ח	ist	rih	111	e	ч	hν